


# PM2.5 FILTER INSPECTION REPORT



**An activated carbon layer is added to the ordinary filter to enhance the filtering effect of the mask.**

**Synthesis:**

two-layer non-woven fabric + 95 meltblown fabric  
+ activated carbon filter + high efficiency filter.

**Application:**

anti-odor, anti-dust, anti-PM2.5  
and other particles;



18001112242



(2018) 国认监认字(244)号



中国认可  
国际互认  
检测  
TESTING  
CNAS L6780

# 检验报告

## TEST REPORT



扫一扫关注我们



扫一扫查询真伪



报告编号

REPORT NO.

国纺委字第 YJ202001776 号

产品名称

NAME OF SAMPLE

PM2.5过滤片

委托单位

CUSTOMER

江苏凌隆企业

检验类别

TEST CATEGORY

委托检验

浙江省轻工业产品质量检验研究院

(浙江省纺织测试研究院)

Zhejiang Light Industrial Products Inspection and Research Institute

国家纺织服装产品质量监督检验中心(浙江)

National Textiles and Garment Quality Supervision Inspection Center(Zhejiang)

浙江省轻工业产品质量检验研究院  
国家纺织服装产品质量监督检验中心（浙江）


# 检验报告

国纺委字第 YJ202001776 号

第 1 页 共 3 页

|   |  |                         |      |
|---|--|-------------------------|------|
| 委托单位名称<br>Name of Customer  | 江苏凌隆企业   | 地址<br>Address           | —    |
| 生产单位<br>Manufacturer  | —  | 地址<br>Address           | —    |
| 样品信息<br>Sample information  | 样品名称 Name of sample: PM2.5 过滤片<br>样品特性 Characteristics: 灰白色<br>商标 Trademark: —<br>规格/号型 Specification/model: —<br>等级 Level: —<br>安全技术类别 Category of safety specification: —<br>样品款号/货号 Art. No.: — |                         |      |
| 以上为客供信息 (Above-mentioned information by Customer-supplied)  |  |                         |      |
| 来样方式<br>The sent way of sample  | 快递   | 样品数量<br>Sample quantity | 1 包  |
| 送检日期<br>Receiving Date of Sample  | 2020-03-12   | 检测类别<br>Test Category   | 委托检验 |
| 判定依据<br>Rating Requirements   | GB/T 32610-2016  |                         |      |
| 检测结论/Test Summary:<br><br>实测结果详见附页。<br><br><p style="text-align: right;">(检验报告专用章) Test Seal</p> <p style="text-align: right;">批准日期/ Date of Approval: 2020-03-12</p> |  |                         |      |
| 备注<br>Remarks   | 1、样品标识未标注防护效果级别，按最低标准要求 D 级判定。<br>2、样品未经预处理。   |                         |      |

签发:  
Approved by





## 声明

- 一、本机构保证检测的公正性、独立性和诚实性，对检测结果负责，对委托方所提供的检测样品保密和保护所有权。
- 二、本报告无签发人员签字或未加盖本机构红色检验检测专用章无效。三、报告涂改无效，复印件未重新加盖本机构红色检验检测专用章无效。四、本报告的检测数据和结果仅对送检样品负责。
- 五、本报告各页均为报告不可分割之部分，使用者单独抽出某些页导致误解或用于其它用途及由此造成的后果，本机构不负相应的法律责任。
- 六、委托方若对本报告有异议，应及时向本机构提出。政府行政管理部下达的指令性任务，被检方对检验结果有异议时，应按政府行政管理部文件规定及国家相关法律、法规规定进行。

## DECLARATION

1. Our organization guarantees impartiality, independence and honesty of inspection, and is responsible for the results of inspection, keeping the samples supplied by the entrusting party confidential and at the same time protecting the ownership of the samples supplied.
2. The test report will be deemed invalid without signatures of the inspector/viewer and authorized personnel, and the red special inspection stamp of our organization.
3. The test report will be invalid if it is altered. Copies of the report are invalid without the red special inspection stamp of our organization.
4. The test results shown in this report is only valid for the tested sample.
5. All the pages of the report are integral parts of the report. Our organization will not be responsible for any misunderstanding or other results caused by using separate page(s) of the report.
6. If there is any dissent of the report, the entrusting party shall notify our organization timely. For the mandatory inspection given by governmental administration departments, and dissent about the sample being tested or test results on the report should be dealt with in accordance with national regulations.

### 浙江省轻工业产品质量检验研究院及附设的检验中心

#### The Affiliated Inspection Centers

浙江省轻工业产品质量检验研究院

联系电话: 0571-85122669

Zhejiang Light Industrial Products Inspection and Research Institute

Telephone: 0571-85122669

国家纺织服装产品质量监督检验中心(浙江)

National Textiles and Garment Quality Supervision Inspection Center(Zhejiang) E-Mail: 1137907994@qq.com 地

址: 浙江省杭州市江干区下沙路300号6号楼

Address: Building No.6, 300 XiaSha Road, Hangzhou, Zhejiang

国家家具产品质量监督检验中心(浙江)

联系电话: 0571-89009556

National Center for Quality Supervision Inspection of Furniture(Zhejiang)

Telephone: 0571-89009556

浙江省室内安全及家具产品质量检验中心

Zhejiang Center of Quality Test for Indoor Safety and Furniture Products E-Mail: 2047699564@qq.com 地址: 浙江省

杭州市余杭区良渚街道经一路1号良渚大学科技园4号楼

Address: Building 4 LiangZhu University Science and Technology Park, No.1,

Jing Yi Rd, Yuhang District, Hangzhou, Zhejiang

浙江省轻工及五金产品质量检验中心

联系电话: 0571-89001107

Zhejiang Center of Quality Test for Light Industry and Hardware Products

Telephone: 0571-89001107

浙江省体育用品质量检验中心

Zhejiang Center of Quality Test for Sports Products

E-Mail: wujtest@126.com

地址: 浙江省杭州市西湖区天目山路 222 号 3 号楼浙江省杭州市余杭区良渚街道经一路  
1 号良渚大学科技园 3 号楼

Address: No. 222 Tianmushan Rd., Hangzhou, Zhejiang

Building 3 LiangZhu University Science and Technology Park, No.1, JingYi Rd., Hangzhou, Zhejiang

国家锁具产品质量监督检验中心 (浙江)

联系电话: 0571-85027738

National Center for Quality Supervision Inspection of Lock(Zhejiang)

Telephone: 0571-85027738

浙江省锁具产品质量检验中心

Zhejiang Center of Quality Test for Lock Products E-Mail: Locktest@126.com 地址: 浙江省杭州市塘苗路 24 号

Address: No. 24 Tangmiao Rd., Hangzhou, Zhejiang

网上业务受理/报告查询: <http://www.zjttj.cn:807>

投诉电话: 0571-85023552

Online Business-Reception/Report Inquires: <http://www.zjttj.cn:807>

Complaint Tel: 0571-85023552



# 检验报告

国纺委字第 YJ202001776 号

第 2 页 共 3 页

| 序号 | 检测项目 | 检测方法                 | 单位 | 标准要求<br>(D 级) | 实测值  | 单项评价 | 结果备注 |
|----|------|----------------------|----|---------------|------|------|------|
| 1  | 呼吸阻力 | GB/T 32610-2016      | Pa | $\leq 175$    | 82.8 | 符合   | —    |
|    | 呼气阻力 |                      | Pa | $\leq 145$    | 39.1 |      |      |
| 2  | 过滤效率 | GB/T 32610-2016 附录 A | %  | $\geq 90$     | 99.6 | 符合   | —    |
| 3  | 防护效果 | GB/T 32610-2016 附录 B | %  | $\geq 65$     | 65.7 | 符合   | —    |



# 检验报告

国纺委字第 YJ202001776 号

第 3 页 共 3 页

样品照片



—以下空白—

